



Declaration of Conformity

Certificate No: 1600

The below listed medical product(s) complies with the
Medical Devices Directive No. 93/42/EEC

The product(s) therefore fulfill all essential requirements for the application of the
CE-conformity mark.

Annex V

Product type: Automatic-inflation electronic sphygmomanometer, portable, arm/wrist
GMDN code: 45617 (UMDNS 16-157)
Product class: Class IIa
Brand name: MICROLIFE

Customers type no.:	Manufacturers type no.:
BP B3 AFIB	BP B3 AFIB (ERP BP3KT1-3N)

Manufacturing plant ONBO ELECTRONIC (SHENZHEN) CO. LTD.

Head office: microlife Corporation
9F, 431, RuiGuang Road, Nei Hu, Taipei 11492, Taiwan, R.O.C.

Manufacturer &
European Subsidiary: Microlife AG
Esenstrasse 139
9443 Widnau
SWITZERLAND

Declare under our sole responsibility that the above product fulfills the essential requirements of the
Directive 93/42/EEC, including all amendments.

Responsible importer: Egli AG Import und Vertrieb
Im Langhang 11
8307 Effretikon
SWITZERLAND

Remarks: Notified Body according to the Medical Devices Directive is TÜV NORD CERT GmbH,
Langemarckstrasse 20, 45141 Essen, Germany. Notified Body ID. No. 0044
Certificate No: 04 235 001845, valid until: 20. December 2023

Place and Date of Issue: Widnau, 28th July 2020
for the Manufacturer:

for the Importer:

microlife
Microlife AG
Esenstrasse 139
9443 Widnau / Switzerland
Phone: +41 71 727 70 50
Teresa Meile
Regulatory Manager