



Konformitätserklärung / Declaration of Conformity

Produktspezifikation / Product details	
Produktebezeichnung / Product name	ActiMaris® NASAL Spray / ActiMaris® NASAL Spray
Type / Type	20 ml
Klassifizierung nach RL 93/42/EWG, Anhang II / Classification according 93/42/EEC, annex II	Klasse I, nach Regel 4 / Class I, Rule 4

Konformitätsbewertung / Assessment details	
Verfahren nach RL 93/42 EWG / Route of directive 93/42/EEC	Anhang VII, annex VII,
Zertifikate / certificates	93/42/ECC

Firmenspezifische Informationen / Company details	
Bevollmächtigter (EU) / Authorised Representative (EU)	Adriatic Health Factory d.o.o. Nikole Tesle 17 / 44000 Sisak / CROATIA

Angewandte Normen / Used standards	
EN ISO 14971: 2019+A11:2021	Medical devices – Risk management
EN ISO 10993-1:2020	Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk management system
EN ISO 15223-1:2021	Symbols for use in the labelling of medical devices

Wir erklären in alleiniger Verantwortung, dass die oben beschriebenen Produkte in der gelieferten Ausführung den Anforderungen der Richtlinie 93/42/EWG und den schweizerischen Medizinproduktegesetzes entsprechen. Die Produkte werden mit dem CE-Kennzeichen versehen.

We declare under sole responsibility that the products described above as delivered are in compliance with directive 93/42/EEC and the medical-device-law of Switzerland. The products are CE marked.



Diese Konformitätserklärung gilt bis / this Declaration of Conformity is valid till: **31.12.2028**

Appenzell, 27.11.2024


ActiMaris AG
Sandgrube 29
9050 Appenzell
Switzerland
www.actimaris.com
Director: Dani Marincic