

2365 Camino Vida Roble Carlsbad, CA 92011 United States

Date: 17-March-2013

Reference: 8352-8900-7-1, 230V Compressor with Nebulizer

Data Sheet

| 8352 Compressor Part Description* | Material |
|-----------------------------------|----------------|
| Plastic Case | Molded ABS |
| Fitting, Exhaust | Molded, Nylon |
| Stem bumper (foot) | Rubber |
| Air filter | Open cell foam |
| Rocker switch | Plastic |

| Temperature: | Humidity: |
|--|---|
| Operation: 10°C to 40°C/ 50°F to 104°F | Operation: 10% to 90%RH non-condensing |
| Storage: -15°C to 50°C / 5°F to 122°F | Storage: 10% to 90%RH non-condensing |
| Transport: -15°C to 70°C/ 5°F to 158°F | Transport: 10% to 90% RH non-condensing |
| | |
| Operating conditions: | Performance Characteristics: |
| Atmospheric pressure: | Air Flow:8 LPM free flow |
| Operation: 700-1060 hPa | 5.5 LPM through a Salter Nebulizer |
| | Pressure: ≧30 PSI Maximum |

| 8900 Nebulizer Part Description | Material |
|---------------------------------|----------------------------------|
| Cup, Nebulizer | Polystyrene |
| Top, Nebulizer | Polypropylene |
| Insert, Nebulizer | Polystyrene |
| Tubing, Aerosol 6" | Ethylene Vinyl Acetate Copolymer |
| T, Anti-drool | K-resin |
| Mouthpiece (Bagged) | High Density Polyethylene |
| Tubing, 3-channel, 7 ft. | Polyvinylchloride |
| Connector, End Piece | Polyvinylchloride |

www.salterlabs.com

Telephone: 760-795-7100 Fax: 760-683-6797



Salter Labs manufactures product from medical grade materials in compliance with Good Manufacturing Practices (GMP's) as listed in 21 C.F.R. (U.S. code of Federal Regulations)

I certify that the above information is accurate to the best of my abilities. Please do not hesitate to contact me if I can be of any further assistance. My email address is mcollins@salterlabs.com and my phone number is 760-795-7102.

Peggy Collins

Regulatory Affairs

Hebby Collins

Salter Labs.

Telephone: 760-795-7100 Fax: 760-683-6797