



2365 Camino Vida Roble
Carlsbad, CA 92011
United States

Date: 16-March-2017

Reference: 8911, Nebulizer Kit with 7 ft. 3-channel tubing.

Part Description	Material
Cup, Nebulizer	Polystyrene
Top, Nebulizer	Polypropylene
Bolo	Polystyrene
Insert, Nebulizer	Green Colorant, MKS
T, Anti-drool	K-resin
Mouthpiece (Bagged)	High Density Polyethylene
Tubing, 3-channel, 7 ft.	Polyvinylchloride
Connector, End Piece	Polyvinylchloride

Salter Labs manufactures product from medical grade materials in compliance with Good Manufacturing Practices (GMP's) as listed in 21 C.F.R. (U.S. code of Federal Regulations).

I certify that the above information is accurate to the best of my abilities.

Please do not hesitate to contact me if I can be of any further assistance. My email address is mcollins@Salterlabs.com and my phone number is 760-795-7102.

Peggy Collins
Regulatory Affairs