

1616 MICRO NASAL CANNULA WITH TRUMPET CONNECTOR - ADULT

LATEX FREE | SINGLE PATIENT USE | DISPOSABLE

| DEVICE MANUFACTURING SPECIFICATIONS | | |
|-------------------------------------|---|--|
| Reference Number | 1616 | |
| Manufacturer | SunMed LLC/Salter Labs | |
| Classification – US | Class I Medical Device | |
| FDA Product Code | CAT – Nasal Oxygen Cannula | |
| Classification – EU | Class IIa | |
| CE Mark/Notified Body | CE 2797/BSI Group | |
| Authorized Representative | Mt Promedt Consulting GmbH | |
| Classification - Canada | Class 2 | |
| EMDN Code | R03010203, Air/Oxygen Nasal Cannula | |
| GMDN Code | 35201 Nasal Cannulas | |
| UMDNS Code | 12799, Cannulae, Nasal Oxygen | |
| Made In | Mexico | |
| Usage | Disposable, Single Patient Multiple Use | |
| Sterile | No | |
| Patient Population | Adult | |
| Packaging | Individually Packaged, 50/Case | |
| Shelf Life | 5 Years | |

Description: Over-the-ear style headset tubing and nasal prongs. Includes oxygen supply tubing with a trumpet connector.

Intended Purpose: A nasal oxygen cannula is a two-pronged device used to administer oxygen to a patient through both nostrils at flows from 0 LPM to 3 LPM.

Area of Use: Hospitals, medical clinics, home, surgical centers, skilled nursing facilities, NICU.

Duration of Use: Discard and replace nasal cannula every 30 days or sooner if the cannula becomes soiled or damaged.

Contraindications: No known.

| Part Number | UOM | GTIN/UDI |
|-------------|------|----------------|
| 1616-4 | Each | 00607411100703 |
| 1616-4-50 | Case | 20607411100707 |
| 1616-7 | Each | 00607411100710 |
| 1616-7-50 | Case | 10607411100717 |

Latex: Salter Labs® does not utilize latex or latex-containing material to manufacture products. To the best of our knowledge, latex does not come in contact with components or finished good during the manufacturing process.

Phthalates: The selected materials are Non-DEHP; DEHP was not intentionally added as part of the manufacturing process.

Mercury-Lead: The selected materials do not contain lead or mercury.

Biocompatible per device classification in ISO 10993. Salter Labs manufactures product from medical grade materials in compliance with Good Manufacturing Practices (GMPs) as listed in 21 C.F.R. (U.S. Code of Federal Regulations).



DEVICE SPECIFICATIONS

| Description | Specification |
|------------------------------|------------------------------------|
| O ₂ Tubing Length | 1' (0.3 m), 4' (1.2 m), 7' (2.1 m) |
| O ₂ Supply Tubing | Clear, 3-Channel Safety |
| O ₂ Flow Rate | 0 LPM to 3 LPM |
| Tubing End Connector | Universal, Trumpet |
| Operating Temperature | 5°C to 40°C |
| Storage Temperature | -20°C to 50°C |

DEVICE MATERIAL

| Plastisol, Reddish Tint |
|------------------------------------|
| Polyvinyl Chloride, 3-Channel |
| Polyvinyl Chloride, Smooth Bore |
| Polyvinyl Chloride |
| Low Density Polyethylene |
| |