

1215 STATIC CONNECTOR

NON-STERILE | DISPOSABLE | SINGLE PATIENT MULTIPLE USE | LATEX FREE | DEHP FREE | BPA FREE

DEVICE MANUFACTURING SPECIFICATIONS		
Reference Number	1215	
Manufacturer	AirLife	
Manufacturer Address	Salter Labs, 60 Spur Road, El Paso, TX 79906	
Country of Origin/Made In	Mexico	
Classification – US/510k #	Class I	
FDA Product Code	BYX - Tubing, Pressure and Accessories	
Classification – Canada	Class II	
Classification – EU	Class IIa, Annex VIII, Rule 2	
CE Mark/Notification Body	CE 2797 / BSI Group	
EU Authorized Representative	Mt Promedt Consulting GmbH	
EMDN Code	R03010204 - Oxygen Administration Tubing	
GMDN Code	12875 – Oxygen Administration Tubing	
UMDNS Code	12875 – Tubing, Oxygen Connecting	
UNSPSC Code	2271715 - Medical oxygen tubing or connectors	
Usage	Disposable, Single Patient Multiple Use	
Patient Population	Infant, Pediatric, Adult	
Area of Use	Hospitals, Pre-Hospital, Home, Surgical Centers, Skilled Nursing Facilities, Medical Clinics	
Sterility	Non-Sterile	
Packaging	50/Poly Bag	
Shelf Life	5 Years	
Disposal Instruction	Used or damaged products must be disposed according to applicable national and international legal directives.	



Intended Purpose: Intended to join two lengths of oxygen supply tubing

Contraindications: No known contraindications..

PART	UOM	GTIN	
1215-0-50	Case	10607411001106	



TECHNICAL SPECIFICATIONS

DESCRIPTION	SPECIFICATION
Flow Rate	≤ 15 LPM
Length	2" (0.05 m)
Tubing End Connector	Tapered Male to Male
Tubing Compatibility	O ₂ Tubing with 5 mm to 7 mm End Connectors
Operating Temperature	+5°C (41°F) to +40°C (104°F)
Storage Temperature	-20°C (-4°F) to +50°C (122°F)

DEVICE MATERIAL

COMPONENT	MATERIAL
Body	High Density Polyethylene (HDPE), Light Green

Biocompatibility: The connectors are biocompatible, meeting cytotoxicity requirements, skin irritation and sensitization requirements per ISO 10993.

The connectors do not contain natural rubber latex, DEHP or $\ensuremath{\mathsf{BPA}}.$

The connectors are REACH compliant and do not contain 3TG conflict minerals.

The connectors do not contain blood derivatives or materials derived from animal or human tissue.