



2365 Camino Vida Roble
Carlsbad, CA 92011
United States

Date: 18 April 2018

Reference: 8501: Replacement filter

Data Sheet

Part Description	Material
Filter	Replacement
Material	Open cell foam
Density	1.9 +/- 0.2 lbs/inch ³
Porosity	100 +/- 10/inch

Salter Labs manufactures product from medical grade materials in compliance with Good Manufacturing Practices (GMP's) as listed in 21 C.F.R. (U.S. code of Federal Regulations)

Please contact Salter Labs Customer Care at customercare@salterlabs.com for further assistance.