

Date: 20-September-2017


Reference: 8906, Nebulizer with pediatric aerosol mask (elastic strap style), 3-channel tubing.

Salter Lab products are not considered chemicals under the OSHA Globally Harmonized System of Classification and Labelling of Chemicals (GHS). The products are designated as “articles” under 29 CFR 1910.1200(c) and are exempt from Safety Data Sheet requirements.

Salter Labs manufactures product from medical grade materials in compliance with Good Manufacturing Practices (GMP's) as listed in 21 C.F.R. (U.S. code of Federal Regulations)

Part Description	Material
Cup, Nebulizer	Polystyrene
Top, Nebulizer	Polypropylene
Bolo	Polystyrene
Insert, Nebulizer	Green Colorant, MKS
Mask, Aerosol (Adult)	Polyvinylchloride
Nose Bar	Aluminum
Elastic Strap	Elastic & Cloth
Grommet, Aerosol	Polypropylene
Tubing, 3-channel	Polyvinylchloride
Connector, End Piece	Polyvinylchloride

Please do not hesitate to contact me if I can be of any further assistance. My email address is mcollins@salterlabs.com and my phone number is 760-795-7102.



Peggy Collins
Regulatory Affairs