



2365 Camino Vida Roble
Carlsbad, CA 92011
United States

Date: March 20, 2015

Reference: 8900, Nebulizer Kit with 7 ft. 3-channel tubing.

Data Sheet


Part Description	Material
Cup, Nebulizer	Polystyrene
Top, Nebulizer	Polypropylene
Insert, Nebulizer	Polystyrene
Tubing, Aerosol 6"	Ethylene Vinyl Acetate Copolymer
T, Anti-drool	K-resin
Mouthpiece (Bagged)	High Density Polyethylene
Tubing, 3-channel, 7 ft.	Polyvinylchloride
Connector, End Piece	Polyvinylchloride

Salter Labs manufactures product from medical grade materials in compliance with Good Manufacturing Practices (GMP's) as listed in 21 C.F.R. (U.S. code of Federal Regulations)

I certify that the above information is accurate to the best of my abilities.

Please do not hesitate to contact me if I can be of any further assistance. My email address is ABrownridge@US.Salterlabs.com and my phone number is 760-795-7102.

Sincerely,



Aurelia Brownridge
Regulatory Associate

Digitally signed by Aurelia Brownridge

Aurelia Brownridge
Regulatory Associate