



Item number 1999902090



Classic

Examination gloves, ABENA Classic, XL, blue, nitrile, powder free

- ✓ Elastic material that adapts to fit the shape of the hand during use
- ✓ Extra soft
- ✓ High level of touch sensitivity
- ✓ Contains no latex proteins



CAT III

Product Description

These elastic nitrile gloves from ABENA provide excellent quality, while maintaining a high standard of hygiene. The gloves adapt to fit the shape of the hand during use, and provide a high level of touch sensitivity. Our disposable gloves are suitable for a wide range of uses such as cleaning tasks, patient care, food handling, laboratory work, medicines handling, chemotherapy, and contact with hormone creams. ABENA's blue nitrile gloves come in a range of sizes and contain no latex proteins, making them safe to use, even for people with latex allergies.



MEDIUM GRIP



EN374-5:2016



374-1:2016/Type B



AQL 1.5 EN EN 420 455

Specifications

Base name	Examination gloves
Brand	ABENA
Sub-brand	Classic
Size	XL
Color	Blue
Properties	Finger textured, latex free, medium grip, rolled cuff, sensitive
Features	Powder free
Single or multiple use	Single use
Material	Nitrile
Technical ingredients	Nitrile NBR, Potassium Hydroxide, Zinc oxide, ZDEC, ZDBC, Sulfur, Titanium Dioxide, Wingstay L, Chlorine
Length/depth	240 mm
Width	115 mm
Thickness	0,08 +/- 0,02 mm
Certifications	CAT III, CE, Food contact materials, MD
CE Category (Personal Protective Equipment)	CAT III
CE Class (Medical Devices)	Class I
Product or test standards	AQL 1.5, EN 374-1:2016 Type B KPT, EN 374-5:2016 VIRUS, EN 420, EN 455
Directives, regulations and acts	MDR (EU) 2017/745, (EU) 2016/425, (EC) No 1935/2004, BEK nr 681 af 25/05/2020, (EC) No 2023/2006
Because of the products composition it has an expiry date from the production date on:	5 years
Storage Instructions	Store dry, clean and at room temperature.
Product Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.
Packaging Disposal Instructions	Can be disposed of with normal household waste sorted according to local regulations.

Instructions for use/application

Examine the gloves for flaws and defects before use.

Packaging data

Unit	Contains	Length	Width	Height	EAN
cII	10 pck	39.7 cm	26 cm	25.4 cm	5703538438103
pck	200 pcs	24 cm	12.3 cm	7.6 cm	5703538438097



Regulation (EU) 2017/745 is a regulation of the European Union on the clinical investigation and sale of medical devices for human use. It repeals Directive 93/42/EEC, which concerns medical devices, and Directive 90/385/EEC, which concerns active implantable medical devices, on 26 May 2021.



The CE mark guarantees that a product is safe to use and complies with all safety precautions. CE stands for Conformité Européenne (European Conformity) and is mainly found on electronic equipment, safety equipment, construction products, and medical devices.



The glass-fork-symbol guarantees that products live up to the demands stated in the European legislation for food contact materials. The symbol is not mandatory but advised on products used for food contact.



Third-party type approval is required for all Category III personal protective equipment (PPE) products. CAT III products will be accompanied by a notified body number for the 3rd party verification of the product's safety.



The product does not contain latex.



The glove has a finger textured surface.



The glove has a medium grip.



The glove is 240 mm long.



The standard specifies the requirements for protective gloves against hazardous chemicals and microorganisms. This part of the standard describes the glove's resistance to bacteria, fungi and viruses.



The standard specifies the requirements for protective gloves against chemicals and microorganisms. Type B has been tested for a permeation time of at least 30 minutes with the chemicals sodium hydroxide 40%, hydrogen peroxide 30% and formaldehyde 37%.



The standard specifies the requirements for protective gloves in relation to general requirements and test methods.



EN 455 consists of four standards that gloves must be tested against in order to be a medical glove. This standard specifies the requirements for disposable medical gloves.

