

TECHNICAL DATA SHEET



ANATOMICALLY SHAPED PADS

SENI SAN CLASSIC



I. PRODUCT DESCRIPTION

Anatomically shaped pads are designed for incontinence protection. Two anatomically shaped absorbent cores are made of cellulose fluff pulp with superabsorbent (in Prima and Normal size – one absorbent core). Contains distributing layer (EDS). Side gathers with elastic yarn along absorbent core are directed outside preventing side leakages. Wetness indicator – black printed inscription that fades, when exposed to wetness. Vapour-permeable outer layer on the whole surface.

II. RAW MATERIALS

Cellulose, superabsorbent, vapour-permeable laminate, hydrophilic-hydrophobic nonwoven, hydrophobic nonwoven, distribution layer (EDS), elastic yarn, ink of indication, hot-melt adhesives, natural green tea extract.

III. DIMENSIONS*

Size	Total length [mm]	Length of absorbent core [mm]		Total width [mm]	Width of absorbent core [mm] (front-crotch-back)
		top	bottom		
Prima	540	-	480	320	240-120-260
Normal	625	-	580	320	240-120-260
Uni	625	370	580	320	240-120-260
Maxi	690	400	610	370	290-170-330
Plus	690	400	610	370	290-170-330

* tolerance of dimensions +/- 10 mm

IV. PROPERTIES

Size	Weight [g] no less than	Absorbency [g] no less than (ISO 11948-1)	Retention [g] average (NAFC)
Prima	44,8	1000	220
Normal	49,8	1200	340
Uni	67,4	1600	450
Maxi	84,7	1900	540
Plus	98,4	2350	740

V. BASIC LOGISTIC DATA

Size	Code	Package [pcs]	Packages per box	Boxes per pallet	Box type
Prima	SE-093-PR30-CL1	30	4	30	carton
Normal	SE-093-NO30-CL1	30	4	28	carton
Uni	SE-093-UN30-CL1	30	4	24	carton
Maxi	SE-093-MA30-CL1	30	4	20	carton
Plus	SE-093-PL30-CL1	30	4	20	carton

Cartons with pads must be placed on wooden pallet 1200x800 mm, covered by PE sheet from the top and PE sheets or cardboard divider sheet from the bottom wrapped together with PE stretch film.

VI. RECOMMENDED STORAGE CONDITIONS

It is recommended to store anatomically shaped pads in a dry and well ventilated warehouse, at stable temperature between 5°C and 35°C and humidity $\leq 70\%$ in their original packaging. The packaging must remain intact without openings in stretch film in order to avoid dust and/or other contaminations.

VII. OTHER

Medical device marked with CE, class I.

Validity: 60 months

Issued by: Monika Landratkiewicz, Technology Specialist

Approved by: Tomasz Przybylski Authorized Proxy /
Deputy Director of Production and Innovation

Date of issue: 17.03.2023

this is legal evidence of all created e-signatures and e-seals

Certificate for a document with an Autenti ID: 1df31ee4-c7fd-4bf3-9394-bd55769fa992
created at: 2023-03-17 14:35 (GMT+01:00)