


PRODUCT TECHNICAL DATASHEET

ITEM REF:	CH-80	REVISION:	10
-----------	-------	-----------	----

LEGAL MANUFACTURER	ICU Medical, Inc. 951 Calle Amanecer, San Clemente, CA 92673, USA						
IMPORTER	ICU Medical B.V. Hofspoor 3, 3994 VZ Houten, The Netherlands						
EU AUTHORISED REPRESENTATIVE (EUAR)	Medical Device Safety Service GmbH (MDSS), Schiffgraben 41, 30175 Hannover, Germany						
ASSEMBLY SITE	ICU Medical de Mexico, S.A. de C., Avenida Cuarza No. 250, Colonia Rancho Santa Clara, Delegacion Manadero Ensenada, Baja California, Mexico 22790						
CLASSIFICATION CODE	GMDN Code: 60539						
INTENDED USE	The ChemoClave is a needle - free CSTD that mechanically prohibits the transfer of environmental contaminants, including bacterial and airborne contaminants into the system, and the escape of drug or vapor concentrations outside the system during drug preparation and administration, thereby minimizing exposure of individuals, healthcare personnel and the environment to hazardous drugs.						
ITEM DESCRIPTION	ChemoClave ® Closed Vial Spike						
PRIMING VOLUME (ml)	0.14	LENGTH (mm)	63	WEIGHT (g)	12	CASE QTY	50
							
ITEM SPECIFIC DATA	Chemical Compatibility		Lipids, Common Chemotherapeutics				
	Luer Compatibility		ISO 80369-7 Compliant male luers > 1,55mm Internal diameter				
	Sterilization and Shelf Life		Radiation; 5-Year Expiration				
	Microbial Ingress and Disinfection Compatibility		Microbial barrier for seven days utilizing a 70% IPA disinfection				
	Extended Use		600 repeat activations				
LIST OF COMPONENTS	1	ENTERPRISE FILTER, 3.0 MICRON		TAPE FILM POLYCARBONATE (NON-FLUID PATH) SILICONE (NON-FLUID PATH) METALIZED NYLON (NON-FLUID PATH)			

PRODUCT TECHNICAL DATASHEET

ITEM REF:

CH-80

REVISION:

10



			PVDF MEMBRANE (NON-FLUID PATH)
	2	CLAVE WHITE SPIKE	SILICONE ACRYLIC POLYESTER (PBT) (NON-FLUID PATH) SILICONE LUBRICANT
MATERIAL COMPLIANCE AND BIOCOMPATIBILITY	This product is made of non-latex, and non-DEHP components. Product has met the requirements of ISO 10993-1 and related standards for biocompatibility.		
PRECAUTIONS	Use aseptic techniques. Single-use only – Do not resterilize. Sterile, Non-Pyrogenic fluid pathway in unopened, undamaged package.		
LABELS AND DIRECTIONS FOR USE	Label and/or DFU contain information for proper use including any warnings or contraindications. Labels are applied on the individual package and on the outside of the cardboard box. Each sales package contains a Direction for Use where applicable.		
PACKING AND PACKAGING	The non-unit packaging is biodegradable as defined by Directive 94/62/EC. Packaging has met the requirements of BS EN ISO 11607 standards. The packaging for this product is Latex Free. This item is blister packed or pouched individually using medical grade materials.		
TRACEABILITY	Lot number provides full traceability of all components and manufacturing processes.		
STORAGE	Store in a dry and clean place. Product should be retained in packaging until ready for use.		
DISPOSAL	The user must dispose of the device according to hospital disposal policy.		
PRODUCTION AND ENVIRONMENT CONTROLS	<ul style="list-style-type: none"> Product is manufactured in an environmentally controlled room. Floor, surfaces, and environment are monitored according to quality procedures. Each component is inspected during acceptance (example: visual; dimensional; functional) to verify conformity with the requirements according to quality procedures. Production and release specific tests are performed according to quality procedures. Limulus amebocyte lysate (LAL) testing is performed on production samples to verify conformity of the endotoxin charge according to quality procedures. Bioburden tests are performed on production samples to verify conformity of the microbial charge according to quality procedures. 		
QUALITY SYSTEM AND PRODUCT CERTIFICATION	Quality System complies to:	ISO 13485:2016	
	Product Certification:	K173477	
	CE Certificate Number:	N/A	
	Notified Body:	N/A	
	Device Classification:	Class II	