


PRODUCT TECHNICAL DATASHEET

ITEM REF: 011-CL2000S REVISION: 07

LEGAL MANUFACTURER	ICU Medical, Inc. 951 Calle Amanecer, San Clemente, CA 92673, USA					
IMPORTER	ICU Medical B.V. Hofspoor 3, 3994 VZ Houten, The Netherlands					
EU AUTHORISED REPRESENTATIVE (EUAR)	Medical Device Safety Service GmbH (MDSS), Schiffgraben 41, 30175 Hannover, Germany					
ASSEMBLY SITE	ICU Medical de Mexico, S.A. de C., Avenida Cuarza No. 250, Colonia Rancho Santa Clara, Delegacion Maneadero Ensenada, Baja California, Mexico 22790					
CLASSIFICATION CODE	GMDN Code: 42750					
INTENDED USE	<p>Neutral-pressure needleless valve-connector</p> <p>A small, sterile, stand-alone, Luer-activated needleless plastic valve intended to mate two related intravenous (IV) line devices [e.g., hypodermic syringe and catheter port or tubing from an IV administration set] and hold them in a secured, sealed, locked position until disconnection, at which point there is minimal fluid flow into or out of the catheter/tubing. It is intended to eliminate the use of needles for IV administration of medications. This is a single-use device.</p>					
ITEM DESCRIPTION	ChemoLock					
PRIMING VOLUME (ml)	0.345	LENGTH (cm)	3.25	WEIGHT (g)	4.35	CASE QTY 50
						
ITEM SPECIFIC DATA	MRI Compatibility	CONTAINS METAL COMPONENTS				
	Chemical Compatibility	LIPIDS, COMMON CHEMOTHERAPEUTICS				
	Sterilization and Shelf Life	Radiation; 5-Year Expiration				
CHEMOLOCK SPECIFIC DATA	Connected Pressure Rating	45 psig / 2327 mmHG				
	Activations	10 Activations				
	Needlefree Connector Compatibility	ChemoLock Port compatible with all known Needlefree Connectors				
	Pressure Rating	ChemoLock Injector = 45 psig / 2327 mmHG; ChemoLock Port = 20 psig / 1034 mmHG				

PRODUCT TECHNICAL DATASHEET

ITEM REF: 011-CL2000S REVISION: 07



	Cytotoxic Drug Compatibility		Solvent and fat emulsion drugs including full strength etoposide and paclitaxel drugs.
	Microbial Ingress and Disinfection Compatibility		Microbial barrier for seven days utilizing a 70% IPA disinfection
SPIROS SPECIFIC DATA	Un-activated Pressure Rating		20 psig / 1034 mmHG
	Activated Pressure Rating		45 psig / 2327 mmHG
	Needlefree Connector Compatibility		Compatible with all known Needlefree Connectors
	Cytotoxic Drug Compatibility		Solvent and fat emulsion drugs including full strength etoposide and paclitaxel drugs.
	Repeat Activations		100 Activations
	Microbial Ingress and Disinfection Compatibility		Microbial barrier for seven days utilizing a 70% IPA disinfection
	Flow Rate		Greater than 200mL/min
LIST OF COMPONENTS	1	POST, FEMALE LUER, SPIROS, 4 CAVITY	POLYCARBONATE
	2	POPPET BODY, CHEMOLOCK, BLUE	POLYCARBONATE
	3	SPRING, CHEMOLOCK, AUTO	STAINLESS STEEL
	4	POPPET, CHEMOLOCK, PE, BLUE	POLYETHYLENE
	5	MAIN SEAL, CHEMOLOCK , RIGID POST	SILICONE
	6	COLLAR, FEMALE LUER, SPIROS	POLYCARBONATE

PRODUCT TECHNICAL DATASHEET

ITEM REF: 011-CL2000S REVISION: 07



	7	FML SEAL, POPPET BODY, CHEMOLOCK 3.0	SILICONE
	8	CAP, BREATHING, MALE	POLYETHYLENE
MATERIAL COMPLIANCE AND BIOCOMPATIBILITY	This product is made of non-latex, and non-DEHP components. Product has met the requirements of ISO 10993-1 for biocompatibility.		
PRECAUTIONS	Use aseptic techniques. Single-use only – Do not resterilize. Sterile, Non-Pyrogenic fluid pathway in unopened, undamaged package.		
LABELS AND DIRECTIONS FOR USE	Label and/or DFU contain information for proper use including any warnings or contraindications. Labels are applied on the individual package and on the outside of the cardboard box. Each sales package contains a Direction for Use where applicable.		
PACKING AND PACKAGING	The non-unit packaging is biodegradable as defined by Directive 94/62/EC. Packaging is according to BS EN ISO 11607. The packaging for this product is Latex Free. This item is blister packed or pouched individually using medical grade materials.		
TRACEABILITY	Lot number provides full traceability of all components and manufacturing processes.		
STORAGE	Store in a dry and clean place. Product should be retained in packaging until ready for use.		
DISPOSAL	The user must dispose of the device according to hospital disposal policy.		
PRODUCTION AND ENVIRONMENT CONTROLS	<ul style="list-style-type: none"> Product is manufactured in an environmentally controlled room. Floor, surfaces, and environment are monitored according to quality procedures. Each component is inspected during acceptance (example: visual; dimensional; functional) to verify conformity with the requirements according to quality procedures. Production and release specific tests are performed according to quality procedures. Limulus amoebocyte lysate (LAL) testing is performed on production samples to verify conformity of the endotoxin charge according to quality procedures. Bioburden tests are performed on production samples to verify conformity of the microbial charge according to quality procedures. 		
QUALITY SYSTEM AND PRODUCT CERTIFICATION	Quality System complies to:	ISO 13485:2016	
	Product Certification:	The product is manufactured in compliance to Council Directive MDD 93/42/EEC as amended.	
	CE Certificate Number:	252.602	
	Notified Body:	NSAI National Standards Authority of Ireland.	
	MDD Device Classification:	Class IIa	