


# PRODUCT TECHNICAL DATASHEET

ITEM REF: 011-0M866-01

REVISION: 09

**icu**medical  
human connections

LEGAL MANUFACTURER	ICU Medical, Inc. 951 Calle Amanecer, San Clemente, CA 92673, USA					
IMPORTER	ICU Medical B.V. Hofspoor 3, 3994 VZ Houten, The Netherlands					
EU AUTHORISED REPRESENTATIVE (EUAR)	Medical Device Safety Service GmbH (MDSS), Schiffgraben 41, 30175 Hannover, Germany					
ASSEMBLY SITE	ICU Medical de Mexico, S.A. de C., Avenida Cuarza No. 250, Colonia Rancho Santa Clara, Delegacion Manadero Ensenada, Baja California, Mexico 22790					
CLASSIFICATION CODE	GMDN Code: 45275					
INTENDED USE	A collection of devices that includes the necessary tubing, pressure transducer(s) and other items [e.g., connectors, stopcock(s), clamps and filters] used for an invasive blood pressure measurement. This set will typically connect directly to the applied invasive catheter and the transducer will provide the electrical signals for display by a patient monitoring system. This device will have physical characteristics appropriate for preserving, to the extent possible, the waveform and fidelity of the measured pressure. This is a single-use device.					
ITEM DESCRIPTION	Dual Port 3ml Squeeze Flush Set Up w/4-Way Stopcock and M/M Adaptor					
PRIMING VOLUME (ml)	0.57	LENGTH (cm)	11.35	WEIGHT (g)	9.19	CASE QTY
						
ITEM SPECIFIC DATA	MRI Compatibility			No metal components		
	Chemical Compatibility			Lipids & Common Chemotherapeutics		
	Luer Compatibility			ISO 80369-7 Compliant.		
	Sterilization and Shelf Life			ETO; 3-Year Expiration		

# PRODUCT TECHNICAL DATASHEET

ITEM REF: 011-0M866-01

REVISION: 09



<b>LIST OF COMPONENTS</b> <i>(Latex and DEHP Free)</i>	1	COVER, VENTED, WHITE, RG	ABS
	2	VENTED WHITE CAP, RG	ABS
	3	SQUEEZE FLUSH DEVICE, 3 ML	POLYCARBONATE POLYPROPYLENE ISOPROPYL ALCOHOL ANHYDROUS USP GRADE SILASTIC
	4	STOPCOCK, 4-WAY LOCKNUT 'THIN WALL', WHITE HA	POLYCARBONATE SILICONE POLYETHYLENE
	5	ADAPTOR, DOUBLE MALE W/NUT	POLYCARBONATE
	6	LUER LOCK NUT, CLEAR	POLYCARBONATE
	7	CAP, LUER LOCK, NON-VENTED, YELLOW	ABS
<b>MATERIAL COMPLIANCE AND BIOCOMPATIBILITY</b>	Product has met the requirements of ISO 10993-1 and related standards for biocompatibility.		
<b>PRECAUTIONS</b>	Use aseptic techniques. Single-use only – Do not resterilize. Sterile, Non-Pyrogenic fluid pathway in unopened, undamaged package.		
<b>LABELS AND DIRECTIONS FOR USE</b>	Label and/or DFU contain information for proper use including any warnings or contraindications. Labels are applied on the individual package and on the outside of the cardboard box. Each sales package contains a Direction for Use where applicable.		
<b>PACKING AND PACKAGING</b>	The non-unit packaging is biodegradable as defined by Directive 94/62/EC. Packaging met the requirements of BS EN ISO 11607. The packaging for this product is Latex Free. This item is blister packed or pouched individually using medical grade materials.		
<b>TRACEABILITY</b>	Lot number provides full traceability of all components and manufacturing processes.		
<b>STORAGE</b>	Store in a dry and clean place. Product should be retained in packaging until ready for use.		
<b>DISPOSAL</b>	The user must dispose of the device according to hospital disposal policy.		
<b>PRODUCTION AND ENVIRONMENT CONTROLS</b>	<ul style="list-style-type: none"> <li>Product is manufactured in an environmentally controlled room. Floor, surfaces, and environment are monitored according to quality procedures.</li> <li>Each component is inspected during acceptance (example: visual; dimensional; functional) to verify conformity with the requirements according to quality procedures.</li> <li>Production and release specific tests are performed according to quality procedures.</li> <li>Limulus amebocyte lysate (LAL) testing is performed on production samples to verify conformity of the endotoxin charge according to quality procedures.</li> </ul>		

## PRODUCT TECHNICAL DATASHEET

ITEM REF: 011-0M866-01

REVISION: 09



	<ul style="list-style-type: none"><li>Bioburden tests are performed on production samples to verify conformity of the microbial charge according to quality procedures.</li></ul>
QUALITY SYSTEM AND PRODUCT CERTIFICATION	<b>Quality System complies to:</b> ISO 13485:2016
	<b>Product Certification:</b> The product is manufactured in compliance to Council Directive MDD 93/42/EEC as amended.
	<b>CE Certificate Number:</b> 252.702
	<b>Notified Body:</b> NSAI National Standards Authority of Ireland.
	<b>MDD Device Classification:</b> Class IIb