


# PRODUCT TECHNICAL DATASHEET

ITEM REF:	011-0P260-01	REVISION:	07
-----------	--------------	-----------	----

LEGAL MANUFACTURER	ICU Medical, Inc. 951 Calle Amanecer, San Clemente, CA 92673, USA						
IMPORTER	ICU Medical B.V. Hofspoor 3, 3994 VZ Houten, The Netherlands						
EU AUTHORISED REPRESENTATIVE (EUAR)	Medical Device Safety Service GmbH (MDSS), Schiffgraben 41, 30175 Hannover, Germany						
ASSEMBLY SITE	ICU Medical de Mexico, S.A. de C., Avenida Cuarza No. 250, Colonia Rancho Santa Clara, Delegacion Manadero Ensenada, Baja California, Mexico 22790						
CLASSIFICATION CODE	GMDN Code: 45275						
INTENDED USE	A collection of devices that includes the necessary tubing, pressure transducer(s) and other items [e.g., connectors, stopcock(s), clamps and filters] used for an invasive blood pressure measurement. This set will typically connect directly to the applied invasive catheter and the transducer will provide the electrical signals for display by a patient monitoring system. This device will have physical characteristics appropriate for preserving, to the extent possible, the waveform and fidelity of the measured pressure. This is a single-use device.						
ITEM DESCRIPTION	1 TP IT w/1 PAED. SAFESET 1 PORT, 48" (121 CM) 30ML/HR						
PRIMING VOLUME (ml)	1.86	LENGTH (cm)	179	WEIGHT (g)	79	CASE QTY	25
							
ITEM SPECIFIC DATA	Chemical Compatibility		Lipids & Common Chemotherapeutics				
	Luer Compatibility		ISO 80369-7 Compliant.				
	Sterilization and Shelf Life		ETO; 3-Year Expiration				
LIST OF COMPONENTS	1	TUBING	PVC				
	2	COHESIVE BELLY BAND	TAPE				
	3	WINGED FEM. LUER	PVC				

# PRODUCT TECHNICAL DATASHEET

ITEM REF: 011-0P260-01 REVISION: 07



	4	LABEL, PATIENT LINE	VARIOUS
	5	LUER CAPS AND COVER	VARIOUS
	6	STOPCOCK, 4 WAY	POLYCARBONATE SILICONE LUBRICANT
	7	CAP, WHITE VENTED	POLYPROPYLENE
	8	COVER, WHITE VENTED	POLYPROPYLENE
	9	SSII PORT TB ASSY	VARIOUS
	10	TRANSDUCER, INTEGR 30 ML WHITE	SILICONE SANTOPRENE SENSOR ADHESIVE HDPE POLYCARBONATE COPPER PVC SILICONE RUBBER
<b>MATERIAL COMPLIANCE AND BIOCOMPATIBILITY</b>		This product is made of non-latex, and non-DEHP components. Product has met the requirements of ISO 10993-1 for biocompatibility.	
<b>PRECAUTIONS</b>		Use aseptic techniques. Single-use only – Do not resterilize. Sterile, Non-Pyrogenic fluid pathway in unopened, undamaged package.	
<b>LABELS AND DIRECTIONS FOR USE</b>		Label and/or DFU contain information for proper use including any warnings or contraindications. Labels are applied on the individual package and on the outside of the cardboard box. Each sales package contains a Direction for Use where applicable.	
<b>PACKING AND PACKAGING</b>		The non-unit packaging is biodegradable as defined by Directive 94/62/EC. Packaging is according to BS EN ISO 11607. The packaging for this product is Latex Free. This item is blister packed or pouched individually using medical grade materials.	
<b>TRACEABILITY</b>		Lot number provides full traceability of all components and manufacturing processes.	
<b>STORAGE</b>		Store in a dry and clean place. Product should be retained in packaging until ready for use.	
<b>DISPOSAL</b>		The user must dispose of the device according to hospital disposal policy.	
<b>PRODUCTION AND ENVIRONMENT CONTROLS</b>		<ul style="list-style-type: none"> <li>Product is manufactured in an environmentally controlled room. Floor, surfaces, and environment are monitored according to quality procedures.</li> <li>Each component is inspected during acceptance (example: visual; dimensional; functional) to verify conformity with the requirements according to quality procedures.</li> <li>Production and release specific tests are performed according to quality procedures.</li> </ul>	

## PRODUCT TECHNICAL DATASHEET

ITEM REF:	011-0P260-01	REVISION:	07
-----------	--------------	-----------	----



QUALITY SYSTEM AND PRODUCT CERTIFICATION	<ul style="list-style-type: none"><li>• Limulus amebocyte lysate (LAL) testing is performed on production samples to verify conformity of the endotoxin charge according to quality procedures.</li><li>• Bioburden tests are performed on production samples to verify conformity of the microbial charge according to quality procedures.</li></ul>	
	Quality System complies to:	ISO 13485:2016
	Product Certification:	The product is manufactured in compliance to Council Directive MDD 93/42/EEC as amended.
	CE Certificate Number:	252.702
	Notified Body:	NSAI National Standards Authority of Ireland.
	MDD Device Classification:	Class IIb